

Clinical Reports

The advantages of the LMA over the tracheal tube or facemask: a meta-analysis

J. Brimacombe MB ChB FRCA

A meta-analysis was performed on randomised prospective trials comparing the laryngeal mask airway (LMA) with other forms of airway management to determine if the LMA offered any advantages over the tracheal tube (TT) or facemask (FM). Of the 858 LMA publications identified to December 1994, 52 met the criteria for the analysis. Thirty-two different issues were tested using Fisher's method for combining the P values. The LMA has 13 advantages over the TT and four over the FM. The LMA had two disadvantages over the TT and one over the FM. There were 12 issues where neither device had an advantage. Advantages over the TT included: increased speed and ease of placement by inexperienced personnel; increased speed of placement by anaesthetists; improved haemodynamic stability at induction and during emergence; minimal increase in intraocular pressure following insertion; reduced anaesthetic requirements for airway tolerance; lower frequency of coughing during emergence; improved oxygen saturation during emergence; and lower incidence of sore throat in adults. Advantages over the FM included: easier placement by inexperienced personnel; improved oxygen saturation; less hand fatigue; and improved operating conditions during minor paediatric otological surgery. Disadvantages over the TT were lower seal pressures and a higher frequency of gastric insufflation. The only disadvantage compared with the FM was that oesophageal reflux was more likely. The importance of these findings in terms of patient outcome could not be determined from the published data.

Une méta-analyse d'études randomisées et prospectives comparant le masque laryné (ML) avec les autres modes de gestion des voies aériennes a été réalisée dans le but de déterminer si le ML présentait des avantages sur le tube trachéal (TT) et le masque facial (MF). Des 858 publications portant sur le ML compilées jusqu'en décembre 1994, 52 ont rencontré les critères de l'analyse. Trente-deux parutions différentes ont été analysées avec la méthode de Fisher pour la combinaison des valeurs de P. Le ML possédait 13 avantages sur le TT et quatre sur le MF. Le ML possédait deux désavantages sur le TT et un sur le MF. Dans douze parutions, aucun des dispositifs n'offrait un seul avantage sur les autres dispositifs. Les avantages du ML sur le TT comprenaient: l'augmentation de la vitesse et de la facilité d'insertion par du personnel inexpérimenté; l'augmentation de la vitesse d'insertion par l'anesthésiste; l'amélioration de la stabilité hémodynamique à l'induction et au réveil; l'amélioration de la saturation en oxygène à la phase de réveil; et une plus faible incidence de maux de gorge. Les avantages sur le MF étaient: la facilité de mise en place par le personnel inexpérimenté; l'amélioration de la saturation en oxygène; une fatigue moindre pour la main; l'amélioration de conditions chirurgicales pendant la chirurgie pédiatrique mineure. Les désavantages sur le TT étaient la baisse de la pression d'étanchéité et une augmentation de la fréquence de l'insufflation gastrique. Le seul désavantage sur le MF était une plus grande susceptibilité au reflux oesophagien. L'importance de ces constatations sur le devenir du patient n'a pu être déterminée à partir des données publiées.

Key words

LARYNGEAL MASK AIRWAY; meta-analysis, risk benefits.

From the Department of Anaesthesia, Cairns Base Hospital, Cairns, Australia 4870.

Accepted for publication 30th June, 1995.

The laryngeal mask airway (LMA) has gained widespread acceptance as a general purpose airway with worldwide usage estimated at over ten million patients by 1993.¹ The popularity of the device for routine use stems from its perceived benefits for the patient and anaesthetist over traditional forms of airway management. Prospective surveys have shown the overall success rate for the technique to be high and the complication rate low,²⁻⁵ but it has been suggested that there is a lack of

randomised prospective trials in peer review journals demonstrating any advantages over the facemask (FM) or tracheal tube (TT),⁶ a view reiterated in a recent review⁷ and accompanying editorial⁸ published in the *Canadian Journal of Anaesthesia*.

The following meta-analysis of randomised prospective trials in peer review journals was performed to determine if the LMA offered any advantages over the TT or FM.

Methods

A literature search was conducted to obtain all publications on the LMA up to December 1994. Three medical databases (48 hours; Medline and Reference Manager Update) were searched for citations containing key words, subject headings and text entries related to the LMA. Additional references were obtained from the indices of anaesthetic journals listed in Index Medicus* and a comprehensive bibliography provided by Intavent, UK.† Prospective randomised studies comparing the LMA with the tracheal tube or facemask were selected for analysis. Only papers and abstracts from peer review journals were included. Data from abstracts published before 1990 or subsequently published as a full paper were excluded.

Data were obtained from the selected papers/abstracts about the type of comparative study (LMA vs TT, LMA vs FM, LMA vs FM vs TT), the study population size and type (adult, paediatric or mixed), the type of surgery (peripheral/superficial or head/neck), the phase of anaesthesia studied (induction, maintenance, emergence, post-operative), ventilation mode, LMA user (anaesthetist or non-anaesthetist), success with the device (considered successful if LMA technique was abandoned in <5% cases)

*Acta Anaesthesiologica Belgica, Acta Anaesthesiologica Scandinavica, Anaesthesia, Anaesthesia and Intensive Care, Anesthesiologie und Réanimation, Anaesthetist, Anesthesiologie, Intensivmedizin, Notfallmedizin, Schmerztherapie, Anesteziologia i Réanimatologia, Anesthesia and Analgesia, Anesthesiology, Annales Françaises D Anesthésie et de Réanimation, British Journal of Anaesthesia, Cahiers d'Anesthésiologie, Canadian Journal of Anaesthesia, Clinical Journal of Pain, European Journal of Anaesthesiology, International Anesthesiology Clinics, Journal of Cardiothoracic and Vascular Anesthesia, Journal of Clinical Anesthesia, Journal of Neurosurgical Anesthesiology, Klinische Anaesthesiologie und Intensivtherapie, Ma Tsui Hsueh Tsa Chi, Japanese Journal of Anesthesiology, Middle East Journal of Anesthesiology, Minerva Anesthesiologica, Regional Anaesthesia, Regional Anesthesia, Resuscitation, Revista Espanola De Anestesiologia y Réanimacion.

†LMA bibliography obtainable from Intavent Research Ltd, Cedar Court, 9-11 Fairmile, Henley-on-Thames, Oxon RG9 2JR, UK

and insertion technique (standard,⁹ other or unknown). All issues addressed by each study were catalogued and their *P* values documented. The criteria for homogeneity were that the study issue and anaesthesia phase were identical. Where the studies were homogenous the *P* values for each issue were pooled and analysed. A null hypothesis was formed that the LMA offered no advantage over the TT or FM for a particular issue. This null hypothesis was tested for each issue using Fisher's method for combining the *P* values.¹⁰ The test statistic for Fisher's method is -2 times the sum of the natural logarithms of the *P* values from each study, and this has a χ^2 distribution with degrees of freedom equal to twice the number of studies. Significance was taken as $P < 0.05$. Only issues addressed by two or more trials were included in the meta-analysis. For convenience, the study issues were divided into six sections: (1) placement; (2) physiology; (3) airway mechanics; (4) airway problems/complications; (5) postoperative aspects and (6) miscellaneous.

Data abstraction

Eight hundred fifty-eight publications on the LMA were catalogued by the author to the end of 1994. This comprised: 160 papers, 132 abstracts, 40 mini-papers (letters containing study data), 61 full case reports, 105 case reports as letters to the editor, 287 other letters, 23 reviews, 15 editorials and 36 other publications, including coincidental use of the LMA in other studies, articles and manuals. The 160 candidate papers and 132 candidate abstracts were examined by two unblinded observers. Where there was a conflict in the observation a third observer was utilised. Two hundred forty papers/abstracts failed to meet the criteria for analysis primarily on the grounds that they were not randomised controlled trials comparing the LMA with other forms of airway management or had not been published in peer review journals. Fifty-two papers/abstracts met the criteria for analysis. This included 37 papers¹¹⁻⁴⁷ and 15 abstracts.⁴⁸⁻⁶² Thirty-five studies compared the LMA with the TT (refs. 11-13, 15, 17-23, 31-35, 37, 38, 40-47, 51, 52, 54-58, 60, 62), 14 studies compared the LMA with the FM (refs. 14, 16, 24-28, 30, 36, 39, 49, 50, 53, 61), one study compared the LMA with the TT and FM⁴⁸ and two studies compared the LMA with a combined TT/FM technique where the TT was replaced by a FM for emergence.^{29,59}

Results

The total study population was 2440 patients and the mean (SD, range) study population size was 47 (30, 10-130). Details about population type, ventilation mode, the LMA user, success with the device and insertion technique are given in Table I. The mean (range) for the

TABLE I Details about population type, ventilation mode, user, success with the device and insertion technique for comparative studies of the laryngeal mask airway (LMA), tracheal tube (TT) and or facemask (FM)

	LMA/FM	LMA/TT	LMA/TT/FM	Total (%)
<i>Population type</i>				
Adult	11	29	1	41 (78.8)
Child	3	6	1	10 (19.2)
Mixed	0	0	1	1 (1.9)
<i>Ventilation mode</i>				
IPPV	3	20	1	24 (46.2)
SV	10	11	2	23 (44.2)
Mixed	1	3	0	4 (7.7)
Unknown	0	1	0	1 (1.9)
<i>User</i>				
Anaesthetist	11	31	3	45 (86.6)
Non-anaesthetist	3	4	0	7 (13.5)
<i>Insertion technique</i>				
Standard	7	5	0	12 (23.1)
Other	0	0	0	0 (0)
Unknown	7	30	3	40 (76.9)
<i>Success rate</i>				
<5% failure	5	15	0	20 (38.5)
≥5% failure	2	5	1	8 (15.4)
Unknown	7	15	2	24 (46.2)

number of issues assessed by each paper/abstract was 4.1 (1-14). A null hypothesis could be constructed for 23 issues between the LMA and TT and nine between the LMA and FM. The LMA had 13 advantages over the TT and four over the FM. The LMA had two disadvantages over the TT and one over the FM. There were 12 issues where neither device had an advantage. Advantages over the TT included: increased speed and ease of placement by inexperienced personnel; increased speed of placement by anaesthetists; improved haemodynamic stability at induction and during emergence; minimal rise in intraocular pressure following insertion; reduced anaesthetic requirements for airway tolerance; lower frequency of coughing during emergence; improved oxygen saturation during emergence; and lower incidence of sore throat in adults. Advantages over the FM included: easier placement by inexperienced personnel; improved oxygen saturation; less hand fatigue; and improved operating conditions during minor paediatric otological surgery. Disadvantages over the TT were lower seal pressures and a higher frequency of gastric insufflation. The only disadvantage compared with the FM was that oesophageal reflux was more likely. The detailed results of the meta-analysis are given in Table II.

Discussion

There were a number of deficits in the analysed studies which limit the findings. Firstly, only 10/52 studies were blinded allowing both user and observer bias. In most circumstances this was not an error of study design, but reflects the difficulties of blinding observational studies between different airway devices. Secondly, in only 12/52 studies was an attempt made to define the LMA insertion technique. It is probable that some techniques may lead to suboptimal positioning and thus influence measured variables.⁶³ Finally, in most studies whilst it was possible to determine if the user was an anaesthetist, their experience with the LMA was usually not defined and there is a probable learning curve with LMA usage.⁶⁴

Despite these shortcomings, this analysis indicates that there is substantial evidence from randomised comparative studies that the LMA has some advantages and disadvantages over the tracheal tube and facemask. Unfortunately the data were not considered to be sufficiently homogenous to make meaningful estimates of the size of the difference between the airway devices. There was also no evidence that any of these differences result in an improvement in patient outcome. However, it is notoriously difficult to demonstrate outcome benefits even with items of anaesthetic equipment that have been universally accepted such as the pulse oximeter.^{65,66} Furthermore, some of the benefits of the LMA cannot be measured in randomised controlled studies particularly in situations where the LMA compliments other forms of airway management or when it is used in unique situations such as airway rescue or as an aid to intubation. Also, where one form of airway management is contraindicated, comparisons between airway devices cannot be made.

There are several areas where the LMA has the potential to benefit patients compared with the TT. The increased speed and reliability of placement by inexperienced personnel suggests a potential role in resuscitation.⁶⁷ The haemodynamic stability at induction and during emergence may be of benefit to patients with cardiovascular disease, and a recent controlled trial comparing normotensive with hypertensive patients supports this concept.⁶⁸ The minimal changes in intraocular pressure may be of benefit to patients with glaucoma.³⁷ The low frequency of coughing during emergence may be beneficial to patients following open eye or ENT surgery where excessive straining is potentially harmful. The LMA may even offer advantages in terms of oxygen saturation during emergence. Postoperatively, the low incidence of sore throat and voice alteration may have advantages for professional voice users and reduce overall morbidity. It should be noted that the advantages of the LMA in terms of sore throat have not been adequately demonstrated in children.

TABLE II Hypotheses tested for laryngeal mask airway versus tracheal tube (LMA vs TT) and facemask (LMA vs FM) using Fisher's method for combining probability scores. *P* values that are underlined confer an advantage to the FM or TT. NS = not significant

	<i>Null hypothesis</i>			
	<i>LMA vs TT</i>		<i>LMA vs FM</i>	
	<i>P value</i>	<i>References</i>	<i>P value</i>	<i>References</i>
<i>Placement</i>				
Ease of placement same for non-anaesthetists	<0.001	11, 15, 34, 38	<0.001	26, 27, 30
Time for placement same for non-anaesthetists	<0.001	11, 15, 34, 38	NS	27, 30
Ease of placement same for anaesthetists	NS	13, 19, 56	-	
Time for placement same for anaesthetists	<0.025	17, 20, 44, 59	NS	39, 59
<i>Physiology</i>				
Pulse rate changes are similar during insertion	<0.001	17, 20, 21, 31, 34, 37, 40, 51	NS	14, 16, 26, 61
Blood pressure changes are similar during insertion	<0.001	17, 20, 21, 31, 34, 37, 40, 51	NS	14, 16, 26, 61
Pulse rate changes are similar during emergence	<0.001	31, 34, 54	-	
Blood pressure changes are similar during emergence	<0.001	31, 34, 54	-	
Intraocular pressure rises are similar during placement	<0.001	12, 19, 20, 31, 37	-	
Frequency of oesophageal reflux is similar	NS	42, 62	<u><0.001</u>	24, 25, 49, 50
Airways are similarly tolerated catecholamine release is similar between devices	NS	31, 40, 43	-	
<i>Mechanical</i>				
Work of breathing is similar between devices	NS	43, 45, 55, 58	-	
Air leak is similar between devices	<u><0.005</u>	35, 48, 56, 57	-	
Gastric insufflation is similar between devices	<u><0.005</u>	18, 35	-	
<i>Airway problems/complications</i>				
Frequency of cough during emergence is similar	<0.001	12, 13, 19, 22, 34, 43	-	
Frequency of laryngospasm NS during emergence is similar	NS	22, 34, 56, 60	-	
Oxygen saturation is similar	<0.01	32, 34, 52, 56, 60	<0.025	14, 26, 29, 39
<i>Postoperative</i>				
Sore throat is similar in adults	<0.05	13, 17, 19, 22, 32, 43, 51	-	
Sore throat is similar in children	NS	14, 47	-	
Voice analysis is similar	<0.001	22, 23	-	
Postoperative surgical pain is similar	NS	32, 43	-	
Postoperative nausea and vomiting is similar	NS	32, 43	-	
<i>Miscellaneous</i>				
Surgical conditions are similar for minor otological surgery in children	-		<0.025	14, 39
Hand fatigue is similar	-		<0.025	26, 59

The main disadvantage of the LMA compared with the TT is that leak and gastric insufflation are more likely. The extent to which this occurs depends on the airway pressure generated and probably also on the precise position of the LMA. Data from very large series have shown that IPPV with the LMA is both safe and effective.³ There were no episodes of gastric dilatation from a series of 11910 LMA anaesthetics.⁶⁹ Leak may be minimised if tidal volume is maintained $<10 \text{ ml} \cdot \text{kg}^{-1}$.⁵⁵ Surprisingly the LMA causes an increased work of breathing compared with the TT even though the larynx, which contributes 25% of total airways resistance, is not bypassed.⁷⁰

The advantage of the LMA over the FM is that it provides a better airway in terms of oxygen saturation, that it does not lead to hand fatigue and is more suitable for IPPV. Operating conditions are also superior during minor paediatric otological surgery. The main disadvantage of the LMA compared with the FM is that reflux is more likely with the LMA. However, this theory remains controversial^{71,72} and dye^{49,73} and oropharyngeal pH studies in both ventilated⁷⁴ and spontaneously breathing patients⁶² have failed to confirm these findings. It has been suggested that the upper oesophageal sphincter may have a role in preventing aspiration.⁷⁵ Data from large audit or epidemiology studies suggest that the in-

idence of aspiration is similar to that for the TT and FM during elective surgery.^{76,77}

This meta-analysis demonstrates that the LMA has several advantages over the TT and FM and a few disadvantages, but does not allow definitive conclusions to be made regarding choice of airway. Further research is needed to determine the importance of these advantages and disadvantages in terms of patient outcome and cost to the health care system and to allow recommendations to be made. Maltby has suggested that randomised controlled trials comparing LMA with the TT and FM for a routine procedure such as arthroscopy would be useful.⁸ However, such studies will require more meticulous design than are demonstrated in many currently available trials. In a recent review, Asai and Morris state that "the true features and role of the LMA will be established only through studies in which the device is used correctly."⁷ It has been suggested that all authors should describe the level of experience of personnel placing the LMA, the method of insertion and where possible fibreoptic scoring should be conducted.^{78,79} Adult studies for routine procedures in which first time insertion rates are less than 90%, overall success rates are less than 95%, or median fibreoptic scoring is less than 3.0, may reflect suboptimal use.⁶⁴

Acknowledgments

I would like to thank A. Berry, D. Mecklem and C. Mann for advice with the manuscript.

References

- 1 *Brimacombe J, Berry A.* The laryngeal mask airway – the first ten years. *Anaesth Intensive Care* 1993; 21: 225–6.
- 2 *Verghese C, Smith TGC, Young E.* Prospective survey of the use of the laryngeal mask airway in 2359 patients. *Anaesthesia* 1993; 48: 58–60.
- 3 *Van Damme E.* Die Kehlkopfmaske in der ambulanten Anästhesie – eine Auswertung von 5000 ambulanten Narikosen. *Anesthesiol Intensivmed Notfallmed Schmerzther* 1994; 29: 284–6.
- 4 *Braun U, Fritz U.* Die Kehlkopfmaske in der Kinderanästhesie. *Anesthesiol Intensivmed Notfallmed Schmerzther* 1994; 29: 286–8.
- 5 *Langer A, Hempel V, Ahlhelm T, Heipertz W.* Die Kehlkopfmaske bei > 1900 allgemeinanästhesien – Erfahrungsbericht. *Anesthesiol Intensivmed Notfallmed Schmerzther* 1993; 28: 156–60.
- 6 *Pennant JH, White PF.* The laryngeal mask airway. Its uses in anesthesiology. *Anesthesiology* 1993; 79: 144–63.
- 7 *Asai T, Morris S.* The laryngeal mask airway: its features, effects and role. *Can J Anaesth* 1994; 41: 930–60.
- 8 *Maltby JR.* The laryngeal mask airway in anaesthesia. *Can J Anaesth* 1994; 41: 888–93.
- 9 *Brain AII.* The intavent laryngeal mask – instruction manual. London: Intavent, 1991.
- 10 *Rosenthal R.* Combining results from independent studies. *Psychol Bull* 1978; 85: 185–93.
- 11 *Pennant JH, Walker MB.* Comparison of the endotracheal tube and laryngeal mask in airway management by paramedical personnel. *Anesth Analg* 1992; 74: 531–4.
- 12 *Holden R, Morsman CDG, Butler J, Clark GS, Hughes DS, Bacon PJ.* Intra-ocular pressure changes using the laryngeal mask airway and tracheal tube. *Anaesthesia* 1991; 46: 922–4.
- 13 *Denry NM, Gadelrab R.* Complications following general anaesthesia for cataract surgery: a comparison of the laryngeal mask airway with tracheal intubation. *J R Soc Med* 1993; 86: 521–2.
- 14 *Johnston DF, Wrigley SR, Robb PJ, Jones HE.* The laryngeal mask airway in paediatric anaesthesia. *Anaesthesia* 1990; 45: 924–7.
- 15 *Davies PRF, Tighe SQM, Greenslade GL, Evans GH.* Laryngeal mask airway and tracheal tube insertion by unskilled personnel. *Lancet* 1990; 336: 977–9.
- 16 *Hickey S, Cameron AE, Asbury AJ.* Cardiovascular response to insertion of Brain's laryngeal mask. *Anaesthesia* 1990; 45: 629–33.
- 17 *Braude N, Clements EAF, Hodges UM, Andrews BP.* The pressor response and laryngeal mask insertion. A comparison with tracheal intubation. *Anaesthesia* 1989; 44: 551–4.
- 18 *Graziotti PJ.* Intermittent positive pressure ventilation through a laryngeal mask airway. Is a nasogastric tube useful? *Anaesthesia* 1992; 47: 1088–9.
- 19 *Akhtar TM, McMurray P, Kerr WJ, Kenny GNC.* A comparison of laryngeal mask airway with tracheal tube for intra-ocular ophthalmic surgery. *Anaesthesia* 1992; 47: 668–71.
- 20 *Wacha MF, White PF, Tychsens L, Stevens JL.* Comparative effects of laryngeal mask airway and endotracheal tube insertion on intraocular pressure in children. *Anesth Analg* 1992; 75: 355–60.
- 21 *Wilson IG, Fell D, Robinson SL, Smith G.* Cardiovascular responses to insertion of the laryngeal mask. *Anaesthesia* 1992; 47: 300–2.
- 22 *Harris TM, Johnston DF, Collins SRC, Heath ML.* A new general anaesthetic technique for use in singers: the Brain laryngeal mask airway versus endotracheal intubation. *J Voice* 1990; 4: 81–5.
- 23 *Lee SK, Hong KH, Choe H, Song HS.* Comparison of the effects of the laryngeal mask airway and endotracheal intubation on vocal function. *Br J Anaesth* 1993; 71: 648–50.
- 24 *Barker P, Langton JA, Murphy PJ, Rowbotham DJ.* Regurgitation of gastric contents during general anaesthesia using the laryngeal mask airway. *Br J Anaesth* 1992; 69: 314–5.

- 25 *Rabey PG, Murphy PJ, Langton JA, Barker P, Rowbotham DJ.* Effect of the laryngeal mask airway on lower oesophageal sphincter pressure in patients during general anaesthesia. *Br J Anaesth* 1992; 69: 346-48.
- 26 *Smith I, White PF.* Use of the laryngeal mask airway as an alternative to a face mask during outpatient arthroscopy. *Anesthesiology* 1992; 77: 850-5.
- 27 *Tolley PM, Watts ADJ, Hickman JA.* Comparison of the use of the laryngeal mask and face mask by inexperienced personnel. *Br J Anaesth* 1992; 69: 320-1.
- 28 *Martin PD, Cyna AM, Hunter WAH, Henry J, Ramayya GP.* Training nursing staff in airway management for resuscitation. A clinical comparison of the facemask and laryngeal mask. *Anaesthesia* 1993; 48: 33-7.
- 29 *Williams PJ, Bailey PM.* Comparison of the reinforced laryngeal mask airway and tracheal intubation for adenotonsillectomy. *Br J Anaesth* 1993; 70: 30-3.
- 30 *Alexander R, Hodgson P, Lomax D, Bullen C.* A comparison of the laryngeal mask airway and Guedel airway, bag and facemask for manual ventilation following formal training. *Anaesthesia* 1993; 48: 231-4.
- 31 *Lamb K, James MFM, Janicki PK.* The laryngeal mask airway for intraocular surgery: effects on intraocular pressure and stress responses. *Br J Anaesth* 1992; 69: 143-7.
- 32 *Swann DG, Spens H, Edwards SA, Chestnut RJ.* Anaesthesia for gynaecological laparoscopy - a comparison between the laryngeal mask airway and tracheal intubation. *Anaesthesia* 1993; 48: 431-4.
- 33 *Wilkins CJ, Cramp PGW, Staples J, Stevens WC.* Comparison of the anesthetic requirement for tolerance of laryngeal mask airway and endotracheal tube. *Anesth Analg* 1992; 75: 794-7.
- 34 *Webster AC, Morley-Forster PK, Dain S, et al.* Anaesthesia for adenotonsillectomy: a comparison between tracheal intubation and the armoured laryngeal mask airway. *Can J Anaesth* 1993; 40: 1171-7.
- 35 *Devitt JH, Wenstone R, Noel AG, O'Donnell MP.* The laryngeal mask airway and positive-pressure ventilation. *Anesthesiology* 1994; 80: 550-5.
- 36 *Epstein RH, Halmi BH.* Oxygen leakage around the laryngeal mask airway during laser treatment of port-wine stains in children. *Anesth Analg* 1994; 78: 486-9.
- 37 *Barclay K, Wall T, Wareham K, Asai T.* Intra-ocular pressure changes in patients with glaucoma. Comparison between the laryngeal mask airway and tracheal tube. *Anaesthesia* 1994; 49: 159-62.
- 38 *Reinhart DJ, Simmons G.* Comparison of placement of the laryngeal mask airway with endotracheal tube by paramedics and respiratory therapists. *Ann Emerg Med* 1994; 24: 260-3.
- 39 *Watcha MF, Garner FT, White PF, Lusk R.* Laryngeal mask airway vs face mask and Guedel airway during pediatric myringotomy. *Arch Otolaryngol Head Neck Surg* 1994; 120: 877-80.
- 40 *Hollande J, Riou B, Guerrero M, Landault C, Viars P.* Comparaison des effets hémodynamiques du masque laryngé et du tube orotrachéal. *Ann Fr Anesth Réanim* 1993; 12: 372-5.
- 41 *Taguchi M, Watanabe S, Asakura N, Inomata S.* End-tidal sevoflurane concentrations for laryngeal mask airway insertion and for tracheal intubation in children. *Anesthesiology* 1994; 81: 628-31.
- 42 *Valentine J, Stakes AF, Bellamy MC.* Reflux during positive pressure ventilation through the laryngeal mask. *Br J Anaesth* 1994; 73: 543-4.
- 43 *Cork RC, Depa RM, Standen JR.* Prospective comparison of use of the laryngeal mask and endotracheal tube for ambulatory surgery. *Anesth Analg* 1994; 79: 719-27.
- 44 *Pace NA, Gajraj NM, Pennant JH, Victory RA, Johnson ER, White PF.* Use of the laryngeal mask airway after oesophageal intubation. *Br J Anaesth* 1994; 73: 688-9.
- 45 *Boisson-Bertrand D, Hannhart B, Rousselot JM, Duvivier C, Quilici N, Peslin R.* Comparative effects of laryngeal mask and tracheal tube on total respiratory resistance in anaesthetised patients. *Anaesthesia* 1994; 49: 846-9.
- 46 *Pennant JH, Pace NA, Gajraj NM.* Role of the laryngeal mask airway in the immobile cervical spine. *J Clin Anesth* 1993; 5: 226-30.
- 47 *Splinter WM, Smallman B, Rhine EJ, Komocar L.* Post-operative sore throat in children and the laryngeal mask airway. *Can J Anaesth* 1994; 41: 1081-3.
- 48 *Barnett R, Gallant B, Fossey S, Finegan B.* Nitrous oxide environmental pollution. A comparison between face mask, laryngeal mask airway and endotracheal tube. *Can J Anaesth* 1992; 39: A151.
- 49 *El Mikatti N, Luthra AD, Healy TEJ, Mortimer AJ.* Gastric regurgitation during general anaesthesia in the supine position with the laryngeal and face mask airways. *Br J Anaesth* 1992; 68: 529P-530P.
- 50 *Owens T, Robertson P, Twomey K, Doyle M, McShane AJ.* Incidence of gastrooesophageal reflux with the laryngeal mask. *Anesthesiology* 1993; 79: A1053.
- 51 *Reinhart DJ.* Comparison of effects of placement of the laryngeal mask airway (LMA) vs endotracheal tube (ETT) on the cardiovascular response. *Anesthesiology* 1993; 79: A1052.
- 52 *Cros AM, Boudey C, Esteben D, Milacic M, Dardel E.* Intubation vs laryngeal mask. Incidence of desaturations and spasms during adenoidectomy. *Anesthesiology* 1993; 79: A1155.
- 53 *Phan Thi H, Ivens D, Khayat Y.* Evolution of the gradient PaCO₂-ETCO₂ during anaesthesia on spontaneous ventilation with face mask versus laryngeal mask. *Anesthesiology* 1993; 79: A37.
- 54 *Joshi GP, Morrison SG, Gajraj NM, Okonkwo N, White*

- PF. Hemodynamic changes during emergence from anesthesia: use of the laryngeal mask airway vs endotracheal tube. *Anesth Analg* 1994; 78: S185.
- 55 Berry A, Verghese C. Changes in pulmonary mechanics during IPPV with the laryngeal mask airway compared to the endotracheal tube. *Anesth Analg* 1994; 78: S38.
- 56 Fiani N, Scandella C, Giolitto N, Prudhomme G, Léon A. Comparison of reinforced laryngeal mask vs endotracheal tube in tonsillectomy. *Anesthesiology* 1994; 81: A491.
- 57 Reinhart DJ, Hansen K, Odesseus K. The laryngeal mask airway and trace gases in the operating room. *Anesthesiology* 1994; 81: A555.
- 58 Joshi GP, Morrison SG, Miciotto CJ, White PF. Evaluation of work of breathing during anesthesia: use of laryngeal mask airway versus tracheal tube. *Anesthesiology* 1994; 81: A1449.
- 59 Waicha MF, Tan TSH, Safavi F, Payne CT, Teufel AE. Comparison of outcome with the use of the laryngeal mask, face mask-oral airway and endotracheal tube in children. *Anesth Analg* 1994; 78: S471.
- 60 Cros AM, Boudey C, Esteben D, Milacic M, Dardel E. Intubation versus masque laryngé - incidence des spasmes et des désaturations en per et postopératoire. *Ann Fr Anesth Reanim* 1994; 12: R87.
- 61 Green D, Ward B, Hughes N. Absence of pressor response following early insertion of laryngeal mask airway after induction with fentanyl and propofol. *Br J Anaesth* 1994; 72: A54.
- 62 Joshi GP, Morrison SG, Okonkwo N, Gajraj NM, Pennant JH, White PF. Continuous hypopharyngeal pH monitoring: use of laryngeal mask airway versus tracheal tube. *Anesthesiology* 1994; 81: A1281.
- 63 Brimacombe J, Berry A. Insertion of the laryngeal mask airway - a prospective study of four techniques. *Anesth Intensive Care* 1993; 21: 89-92.
- 64 Brimacombe J. Analysis of 1500 laryngeal mask uses by one anaesthetist in adults undergoing routine anaesthesia. *Anaesthesia* (in press).
- 65 Moller JT, Pedersen T, Rasmussen LS, et al. Randomized evaluation of pulse oximetry in 20,802 patients: I. Design, demography, pulse oximetry failure rate, and overall complication rate. *Anesthesiology* 1993; 78: 436-44.
- 66 Moller JT, Johannessen NW, Espersen K, et al. Randomised evaluation of pulse oximetry in 20,802 patients: II. Perioperative events and postoperative complications. *Anesthesiology* 1993; 78: 445-53.
- 67 Baskett PJF. The laryngeal mask in resuscitation (Editorial). *Resuscitation* 1994; 28: 93-5.
- 68 Fujii Y, Tanaka H, Toyooka H. Circulatory responses to laryngeal mask airway insertion or tracheal intubation in normotensive and hypertensive patients. *Can J Anaesth* 1995; 42: 32-6.
- 69 Verghese C, Brimacombe J. Survey of laryngeal mask airway usage in 11910 patients. Safety and efficacy for conventional and non-conventional usage. *Anesth Analg* (in press).
- 70 Tully A, Brancatisano A, Loring SH, Engel LA. Influence of posterior cricoarytenoid muscle activity on pressure-flow relationship of the larynx. *J Appl Physiol* 1991; 70: 2252-8.
- 71 Vanner RG. Regurgitation and the laryngeal mask airway (Letter). *Br J Anaesth* 1993; 70: 380-1.
- 72 Brimacombe J, Berry A. Aspiration pneumonitis and the laryngeal mask airway (Letter). *Anesth Analg* 1994; 78: 816.
- 73 Akhtar TM, Street MK. Risk of aspiration with the laryngeal mask. *Br J Anaesth* 1994; 72: 447-50.
- 74 Lefort P, Visseaux H, Gabriel R, Palot M, Pire JC. Utilisation du masque laryngé pour le coelioscopie. *Ann Fr Anesth Reanim* 1993; 12: R231.
- 75 Vanner RG. Gastro-oesophageal reflux and regurgitation during general anaesthesia for termination of pregnancy. *Int J Obstet Anesth* 1992; 1: 123-8.
- 76 Brimacombe J, Berry A. The incidence of aspiration associated with the laryngeal mask airway - a meta-analysis of published literature. *J Clin Anesth* 1995; 7: 297-305.
- 77 Haden RM, Pinnock CA, Scott PV. Incidence of aspiration with the laryngeal mask airway (Letter). *Br J Anaesth* 1994; 72: 496.
- 78 Brimacombe J, Berry A. Research and the laryngeal mask airway (Letter). *Anesthesiology* 1993; 79: 411-2.
- 79 Brimacombe J, Berry A. A proposed fiber-optic scoring system to standardize the assessment of laryngeal mask airway position (Letter). *Anesth Analg* 1993; 76: 457.